AMENDMENT NO. 1 DECEMBER 2016
TO
IS 6190 : 1971 SPECIFICATION FOR ANAESTHETIC FACE MASKS

(*Page* 3, *clause* **5.1**) **—** Substitute the following for the existing clause:

"**5.1**The face mask shall have a female conical fitting conforming to IS 3234 (Part 2) : 1995 'Conical fittings with a 6 percent (LUER) taper for syringes, needles and certain other medical equipment: Part 2 Lock fitting (*second revision*)'."

(*Page* 3, *footnote* *) **—** Delete.

(MHD 11)

Publication Unit, BIS, New Delhi, India